

STANDARD F

Legionella Ag FIA SD BIOSENSOR


STANDARD F Legionella Ag FIA test system (Analyzer + Test device) detects Legionella pneumophila serogroup 1, 3, 5, 6 and 8 antigens via easily collectible urine sample. Without further sample processing, STANDARD F Legionella Ag FIA performs high sensitivity, and the test is not affected by Rheumatoid factor.



Specification

Intended use	Qualitative detection of L. pneumophila serogroup 1, 3, 5, 6 and 8 antigens in urine specimens.
Specimen volume	100 µl
Specimen type	Urine
Measurement range	Quantitative result with cut-off index (COI)
Testing time	15 mins (Early detection available)
Storage condition	2 - 30°C / 36 - 86°F

STANDARD TEST MODE



READ ONLY MODE



TEST PERFORMANCE

Reference	Sensitivity	Specificity
Flourescent Immunoassay	94.9% (37/39)	100% (40/40)

ORDERING INFORMATION

Category	Product	Pack Size
Respiratory	STANDARD F Legionella Ag FIA	25 Test
	STANDARD F Legionella Ag Control	Pos x 10 / Neg x 10