

STANDARD F


Zika Ag FIA SD BIOSENSOR

STANDARD F Zika Ag qualitatively analyzes Zika virus antigen in serum, plasma and whole blood using fluorescent immunoassay. Screening test results can be obtained within 15 minutes with high sensitivity & specificity.



Specification	
Intended use	Qualitative detection of Zika virus antigen
Specimen Type	Whole blood, Serum, Plasma
Specimen Volume	100 µl
Testing Time	5-15 mins.
Storage Conditions	2-30°C / 36-86°F

STANDARD TEST MODE



READ ONLY MODE



TEST PERFORMANCE

Reference	Sensitivity	Specificity
RT-PCR	96.3% (52/54)	97% (150/154)

ORDERING INFORMATION

Category	Product	Pack Size
Vector Borne	STANDARD F Zika Ag FIA	25 Test